Excitation function of (p, α) nuclear reaction on enriched ⁶⁷Zn: Possibility of production of ⁶⁴Cu at low energy cyclotron

By F. Szelecsényi¹*, Z. Kovács¹, K. Nagatsu², M-R. Zhang² and K. Suzuki²

Received; accepted

Medical radioisotope: ⁶⁴Cu / Proton induced reaction / Excitation function / Integral yield / ⁶¹Cu and ⁶⁷Cu contamination levels

Abstract. The potential for production of the medically relevant ⁶⁴Cu has been investigated by proton irradiation of highly enriched ⁶⁷Zn targets. The excitation function of the 67 Zn(p, α) 64 Cu a nuclear reaction was measured by the stacked-foil technique up to 30 MeV. The prediction of the TALYS code was also compared to the measured cross section results. Based on the improved database of the 67 Zn(p, α) 64 Cu reaction, thick target yield as a function of energy was also deduced. Production possibility of ⁶⁴Cu is discussed in detail, employing different energy proton beams and with regards to the ⁶¹Cu and ⁶⁷Cu contamination levels as a function of the target enrichment level. By using 1 µA beam intensity, 6.3505 h irradiation time and enriched ⁶⁷Zn target (⁶⁴Zn≤0.5%, ⁶⁶Zn≤9%, ⁶⁷Zn≥80%, ⁶⁸Zn≤10% and ⁷⁰Zn≤0.5%), the expected EOB (End Of bombardment) yields are 43.66, 88.80 and 156.14 MBq/µA at 12, 15 and 18 MeV proton energies, respectively. Application timeframes were also deduced where the total radio-copper contamination level remains below 1%.

1. Introduction

Among the copper radioisotopes ⁶⁴Cu is of special interest to nuclear medicine as it can be employed both for medical imaging (via Positron Emission Tomography: PET) and for targeted radio-immunotherapy of tumours. Its relative long half-life ($T_{1/2}$ =12.701 h [1]; decay scheme: EC(43.7%) [1], β^{+} (38.5%) [1], β^{+} (17.8%) [2]) not only makes possible performing investigations with ⁶⁴Cu labelled compounds over several days, but - as an additional benefit – it is short enough to limit the patient's exposure during these studies [cf. 3-5].

*Author for correspondence (E-mail: szele@atomki.hu)

¹ Cyclotron Application Department, Institute for Nuclear Research, Hungarian Academy of Sciences, Bem tér 18/c, Debrecen H-4026, Hungary

² Molecular Imaging Center, National Institute of Radiological Sciences, 4-9-1 Anagawa, Inage-ku-Chiba, 263-8555, Japan

Several research centres have already investigated its production routes via proton (and deuteron) induced nuclear reactions on highly enriched 64 Ni (natural isotopic composition of 64 Ni: 0.926%) target [*cf.* 6-9]. Nowadays, the 64 Ni(p, n) 64 Cu process is used in practice providing the purest form of 64 Cu (both radionuclidic and chemical) in very high yield.

Due to the relative high price of the enriched nickel material, however, alternative 64 Cu production methods were also studied worldwide in recent years. As a result of these detailed studies, proton and deuteron induced reactions on highly enriched 64 Zn, 66 Zn and 68 Zn targets have been suggested as candidates for this purpose. The investigated nuclear reactions were as follows: 64 Zn(d, 2p) 64 Cu, 66 Zn(d, α) 64 Cu, 66 Zn(d, α) 64 Cu, 68 Zn(p, α n) 64 Cu and 68 Zn(d, α 2n) 64 Cu [cf. 10-13]. Compilation and evaluation of these other production routes of 64 Cu was already the subject of an IAEA's Coordinated Research Project [14]. Furthermore, Aslam $et\ al.$ [15] have also published a comprehensive evaluation of these production related cross section data.

From the point of view of a lower energy 'biomedical' cyclotron (around 12-18 and 6-10 MeV maximum proton and deuteron energy, respectively), the starting energies of 66 Zn(p, 2pn) 64 Cu (Q= -18.83 MeV), 68 Zn(p, αn) 64 Cu (Q= -7.79 MeV) and 68 Zn(d, $\alpha 2n$) 64 Cu (Q= -10.014 MeV) nuclear reactions are too high to use these processes for practical purposes. Although the above values for the 64 Zn(d, 2p) 64 Cu (Q= -2.02 MeV) and 66 Zn(d, α) 64 Cu (Q= +7.24 MeV) seems to be acceptable, but unfortunately their reported excitation function curves reach their maximum beyond the available deuteron energy ranges [11,13]. Additionally, the limited number of higher energy deuteron accelerators in research centres could also limit the widespread application of the 'deuteron ways'.

Surveying the information on cross section data of all 64 Cu producing Zn+p and Zn+d reactions, it was found one process that could be useful even at low energies, namely the 67 Zn(p, α) 64 Cu (Q= +2.407 MeV) reaction [16]. The excitation function of this reaction was studied by only one author in the past, probably because of the low isotopic composition of 67 Zn (4.1%) in natural zinc [17]. It should be noted, however, that in spite of the fact that highly enriched 67 Zn material is relative expensive (compared to 68 Zn) the 67 Zn+p reactions are used for routine 67,66 Ga production at low energy cyclotrons [cf. 18,19]. Although numerous cross section measurements can be found in the literature for 67 Zn+p reactions, it is surprising that the database of those processes that produce copper radioisotopes is rather scanty [16].

To evaluate the practical production circumstances of 64 Cu (*i.e.* production energy range, activation time, contamination level(s), time-frame of use etc.) at a biomedical cyclotron via the 67 Zn(p, α) 64 Cu reaction, it is important to have reliable cross section databases of all Zn+p reactions that form 60,61,62,64,67 Cu radioisotopes below 18 MeV. The 64 Cu product certainly has radio-copper contamination(s) if a target with lower enrichment level (*i.e.* <100% of 67 Zn) is activated. From the point of view of 64 Cu, the presence of 61 Cu(T_{1/2} = 3.333 h) and 67 Cu(T_{1/2} = 61.83 h) [1] at EOB

means the major radio-copper contaminations. Although other copper radioisotopes can be also formed below 18 MeV, their half lives are too short (for example: 60 Cu($T_{1/2}$ = 23.7 min) or 62 Cu($T_{1/2} = 9.673$ min)) to cause any impurity problem at the time of the practical application. The amount of ⁶¹Cu (shorter lived contamination) can prolong the waiting period (cooling time) before the medical applications. Similarly, the presence of ⁶⁷Cu at EOB may limit the length of the application period of the labelled compounds. Thanks to the available detailed excitation function studies, the cross section database of 64 Zn(p, α) 61 Cu (Q= +0.844 MeV), 68 Zn(p, 2p) 67 Cu (Q= -9.977 MeV) and 70 Zn(p, α) 67 Cu (Q= +0.262 MeV) reactions seem to be well measured and can be used for calculating the thick target yields of the above reactions with the required precision. The only exemption is the 67 Zn $(p, \alpha)^{64}$ Cu reaction, that - as it was mentioned above - has only one dataset.

To improve the database of the 67 Zn(p, α) 64 Cu nuclear reaction, we decided to re-measure its excitation function curve up to 30 MeV and compare the new dataset not only to the available experimental results [17], but also with the predicted ones calculated with the 4^{th} version of TALYS code from the TENDL-2012 online library [20]. Based on the evaluated cross section databases of the contributing reactions, we could calculate the 64 Cu EOB yields as function of energy and irradiation time. Additionally, it becomes also possible to estimate the EOB radio-copper contamination levels of the final product as a function of irradiation time and target enrichment level.

We discuss here in detail the actual production possibility of ⁶⁴Cu via irradiations with three different energy proton beams (*i.e.* 12, 15 and 18 MeV).

2. Experimental

2.1 Targets

Cross sections for ⁶⁴Cu were measured via the activation technique by bombarding the samples in a 'stacked-form' arrangement. Five pieces of highly enriched ⁶⁷Zn targets (64Zn: 1.44%; 66Zn: 2.2%; 67Zn: 91.5±0.5%; 68Zn: 4.78%; 70Zn: 0.08%; (molar fraction)) were prepared via electrodeposition. The preparation method was similar to those given in [18]. The enriched material was supplied by v/o Technabexport, Moscow, Russia. Commercially available thin natural Ni foils (thickness: 8.90 mg/cm²; Goodfellow Metals, UK) served as target backing material. Other thin metal foils (Ti (9.08 mg/cm²) and Cu (8.92 mg/cm²)) which were also used during the experiments (i.e. for energy degradation and beam intensity monitoring) were purchased also from the above source. The diameter of the electroplated ⁶⁷Zn targets was 10 mm while their thicknesses varied between 7.85 and 12.70 mg/cm². The targets and foils used were individually weighed before the activations to evaluate their actual thickness.

2.2 Irradiations and beam current measurement

Two irradiations were done at the AVF-930 isochronous cyclotron of NIRS, Chiba with the same primary proton energy (30.6±0.4 MeV). The energies were determined by magnetic deflection. The same five ⁶⁷Zn samples were activated in both cases in a special target holder that served also as a Faraday-cup for charge measurement. The time difference between the two activations was almost 6 days. This way the ⁶⁴Cu activity produced during the 1st experiment could decay out until the beginning of the second activation. The ⁶⁷Zn samples were interspersed with copper energy degrader (and monitor) foils. The number of copper foils was different in the stacks to get ⁶⁴Cu cross sections at different energies. To prevent the contamination of the targets with cold and radioactive copper isotopes originating from the Cu foils, Ti foils were placed in front of each Zn samples (64 Cu is also formed via the 65 Cu(p, pn) (Q= -9.91 MeV) and 65 Cu(p, d) (Q= -7.69 MeV) reactions). Naturally, special care was taken during the assembling and disassembling of the stacks to prevent additional copper contamination of the target samples.




Fig.1. Excitation functions of $^{\text{nat}}\text{Cu+}p$ monitor reactions used in this study for beam monitoring.

Both activations lasted for 2 hours with the same beam current of 100 nA. An Al collimator (thickness: 3 cm) was used to get a well collimated beam ($\emptyset = 4$ mm) on the surface of the foils. Similar to our previous work on $^{68}\text{Zn}+p$ reaction [21], the values of the accumulated charges were also compared to the calculated ones, obtained via the IAEA's recommended $^{\text{nat}}\text{Cu}+p$ monitor reaction database [22]. In the case of the first experiment both the $^{\text{nat}}\text{Cu}(p,xn)^{62}\text{Zn}$ and $^{\text{nat}}\text{Cu}(p,n)^{63}\text{Zn}$ processes were used, while during the 2^{nd} activation of the foils only the $^{\text{nat}}\text{Cu}(p,n)^{63}\text{Zn}$ reaction proved useful.

In Fig.1 our measured cross sections of ⁶²Zn and ⁶³Zn, are compared to the recommended values. Beam fluxes determined by direct integration agreed well (within 6%) with the monitor reaction results.

2.3 Activity measurement

The irradiated samples and monitor foils were measured non-destructively using an HPGe detector (EGC15-185-R, 76 cm³, Eurisys Measures, France) connected to an MCA win2000 data acquisition system. The energy calibration of the detector and the determination of its counting efficiency were done using standard gamma-ray point sources (²²Na, ⁵⁴Mn, ⁵⁷Co, ⁶⁰Co and ¹³⁷Cs) supplied by Isotope Products Laboratories (Burbank, California, USA). The uncertainty of these calibration sources were $\pm 3\%$. All foils were measured at distances of ≥ 20 cm (from the detector surface), minimizing this way the random pile-ups and summations, and reducing dead-time to less than 3 %. Additionally, from this distance the requirement for 'point-like sources' could be also assured. (The activities produced in the samples were concentrated in the center of the foils within a small spot with a diameter of 4 mm.)

Since the enriched 67 Zn material was electroplated onto nat Ni material, 64 Cu was produced also via the 64 Ni(p, n) 64 Cu reaction in the activated samples. Based on the evaluated database of the 64 Ni(p, n) 64 Cu reaction [22] and taking into account the thickness of the nat Ni backings, the abundance of 64 Ni (0.926%) in nat Ni, and the irradiation circumstances, we could estimate the amount of 64 Cu activity produced in each target backings during the different activations. We compared these estimated activity values with the actual 64 Cu ones measured from the samples, and the found them negligible.

The decay characteristics (*i.e.* dominant gamma-ray energies and their branching ratios) of the measured Ga, Zn and Cu radioisotopes were taken from the Nudat 2.6 database [1].

Since only one γ -ray is emitted in the decay of ⁶⁴Cu with low intensity, the measurement of the ⁶⁴Cu activity via this line is not so straightforward (1345.8 keV (I γ = 0.48%)). There is however, a 'counting window' (between 1.5 and 2.5 half lives after EOB) when ⁶⁴Cu can be assayed easily [12]. According to the conclusions reported in [23], the possibility to measure the ⁶⁴Cu activity using the annihilation-peak without chemical separation (decay-curve analysis) was ruled out.

2.4 Calculation of reaction cross sections and their uncertainties

The cross sections were calculated by applying the well-known activation formula. Corrections were made for decay losses during and after bombardment, as well as during counting. No correction was required for the recoil effect as the Ni backings served also as catcher foils. Since ⁶⁷Zn had only 91.5% abundance in the target material, the obtained cross section values had to be normalized to 100% enrichment level.

The energy degradation along the stacks and the effective particle energy in the middle of each target foil were calculated according to the polynomial approximation of Andersen and Ziegler [24].

Table 1. Measured cross sections for the 67 Zn(p, α) 64 Cu nuclear reaction.

Energy	Cross section of
(MeV)	67 Zn(p , α) 64 Cu (mb)
7.9±1.7	7.3±1.3
9.6±1.6	16.7±2.2
11.1±1.6	24.9±2.8
15.8±1.4	26.3±2.9
18.1±1.3	19.1±2.4
19.4±1.2	15.7±1.8
21.0±1.2	13.2±1.7
21.4±1.1	11.2±1.5
24.5±1.0	7.3±1.2
28.5±0.9	5.4±1.1




Fig.2. Excitation function of the 67 Zn(p, α) 64 Cu nuclear reaction.

Due to the straggling effect and foil thickness uncertainties the initial uncertainty in the proton beam energy increased along the stacks and reached ± 1.3 and ± 1.7 MeV in the last foils of the 1^{st} and 2^{nd} experiments, respectively,

The estimation of standard uncertainty on the cross section values supposing equal sensitivities for the different parameters was performed as described in [25]. The following individual uncertainties were included in the propagated error: foil thickness or the number of target nuclei, including target non-uniformities (2% but 5% for the electroplated ⁶⁷Zn); incident particle intensity (4–6%); detection efficiency (3–5%), depending on the energy of the gammaphoton; determination of the peak areas, including statistical errors (3–15%); abundance of the gamma rays analysed (1-2%). Only the linearly contributing independent uncertainty sources were used and the non-linear sources were omitted in the calculation. The uncertainties of time information (irradiation, cooling and measuring time and halflife) were neglected since their contributions are not significant in this case. The resulting total average uncertainties amount to 10-12% for monitors and 13-17% for ⁶⁴Cu cross sections, obtained as square root of the sum of squares of the contributing sources.




Fig.3. ⁶⁴Cu activity as a function of bombarding proton energy.

3. Results and Discussion

3.1 Cross sections of the 67 Zn(p, α) 64 Cu nuclear reaction

The excitation function for the 67 Zn(p, α) 64 Cu nuclear reaction is shown by closed circles in Fig.2. Table 1 contains the numerical values of the measured cross sections (10 data points) and their uncertainties. To check the reliability of our measurement we have compared the new cross sections to the available theoretical and experimental data. As it was mentioned earlier, only Levkovskij [17] measured this reaction before our study. Unfortunately, the details of his experimental circumstances were not available in the original work. It is supposed that the 1345.8 keV gamma line was employed by him for cross section measurement. Recently, Takács et al. [26] have shown that all cross section data for proton induced processes reported by Levkovskij [17] should be decreased by 20% based on the presently accepted value of the monitor reaction ($^{\text{nat}}\text{Mo}(p,$ $(x)^{96m}$ Tc) that was used by Levkovskij in his original work. Therefore, for comparison we had to re-calculate his values for the 67 Zn(p, α) 64 Cu nuclear reaction. These 'updated' data are reproduced in Fig.2. The predicted excitation function curve calculated via the TALYS code [20] is also added to Fig.2 up to 35 MeV.

The two experimental data sets show good agreement with each other not only in the cross section magnitude but they gave the same energy for the position of the peak of the excitation function curve as well ($\sigma_{max} = 30$ mb at about 14 MeV). Although the theoretical calculation predicts the same peak position, it overestimates the cross section values over the whole onset part of the excitation function curve. In contrast to this overestimation, the slope part of the curve is in good agreement with the experimental results (beyond 17 MeV).

For thick target yield calculations, therefore, we decided to use the 'eye-fitted' excitation function curve (see Fig.2) of the 67 Zn(p, α) 64 Cu reaction, based on the data of this

work and the values of Levkovskij [17], instead of the predicted values.




Fig.4. ⁶¹Cu/⁶⁴Cu and ⁶⁷Cu/⁶⁴Cu activity ratio and ⁶⁴Cu yield as a function cooling time using 12 MeV incident proton energy.

3.2 Production possibility of 64 Cu via the 67 Zn(p, α) 64 Cu reaction

The available ⁶⁴Cu activities as a function of the bombarding energy are reproduced in Fig.3 for enriched ⁶⁷Zn target. One half of the half-life of ⁶⁴Cu irradiation time was used during this calculation. The 6.3505 h activation time seems to be acceptable at biomedical cyclotrons having busy production agenda. During this time almost 30% of the saturation activity of ⁶⁴Cu could be produced. As it can be seen in Fig.3., the yield increases with the increasing energy. It is interesting to mention that the thick target yields of ⁶⁴Cu calculated via the theoretical excitation function curve overestimate the yields based on the experimental data by a factor of 2. As it is expected, the natural zinc target produces the lowest EOB activity (9.25 MBq/µA, (0.25 mCi/µA)) at 20 MeV. It is worth considering that in the case of ^{nat}Zn targets the 68 Zn(p, αn) 64 Cu reaction also contributes to the formation of ⁶⁴Cu above 10 MeV. The low yield and the long cooling time due to the co-formed ⁶¹Cu (huge amount via the 64 Zn(p, α) 61 Cu reaction) disqualifies the nat Zn target to use for practical purposes. For example, at 15 MeV more than 56 h cooling time would be necessary to reach <1% 61Cu/64Cu activity ratio; resulting this way only 370 kBq/µA (0.01 mCi/µA) practical yield. On the other hand, the longer lived radio-copper contaminant, the ⁶⁷Cu, (via the ⁶⁸Zn(p, 2p)⁶⁷Cu and ⁷⁰Zn(p, α)⁶⁷Cu reactions) [22, 27, 28] would decrease the length of the application time (below 11 h; <1% ⁶⁷Cu contamination) of the final product. Note, that the presence of ⁶¹Cu and ⁶⁷Cu in the final product has not only influence on the quality of the PET measurement, but also delivers unnecessary dose to the patient. Therefore, the production method selected must keep the contamination levels as low as possible (usually below 1%.)

A zinc target with $100\%^{67}$ Zn enrichment level provides not only the highest yield (225.59 MBq/ μ A (6.07 mCi/ μ A) at 20 MeV)) but the product will be also radio-copper contamination-free as well. Although above 16.37 MeV the formation of 62 Cu ($T_{1/2} = 9.74$ min) becomes possible (via the 67 Zn(p, o2n) 62 Cu reaction) but - as it was discussed earlier - it decays out from the final product during the processing time.

It is obvious that due to high price of 100% enriched ⁶⁷Zn, target materials with lower ⁶⁷Zn enrichment levels are the real options for practical purposes.




Fig.5. ⁶¹Cu/⁶⁴Cu and ⁶⁷Cu/⁶⁴Cu activity ratio and ⁶⁴Cu activity as a function cooling time using 15 MeV incident proton energy.

As it can be calculated (Fig.4), acceptable yields (\geq 37 MBq/µA) can be achieved even at 12 MeV using a target containing more than 80% ⁶⁷Zn. Depending, however on the actual composition of the target, different ⁶¹Cu/⁶⁴Cu and ⁶⁷Cu/⁶⁴Cu activity ratios could be observed at EOB. In analyzing the excitation functions of those Zn+p reactions that form ⁶¹Cu and ⁶⁷Cu below 18 MeV, the following can be concluded concerning the purity level of the final ⁶⁴Cu product: a) the amount of ⁶⁴Zn and ⁷⁰Zn in the actual target has the major influence on the purity of the final product [14,27,28]. b) The amount of ⁶⁶Zn and ⁶⁷Zn in the target matrix is indifferent from the point of view of ^{61,67}Cu formation.

Below we discuss in details the available ⁶⁴Cu yields and the ⁶¹Cu/⁶⁴Cu and ⁶⁷Cu/⁶⁴Cu activity ratios as a function of the cooling time at 12, 15 and 18 MeV irradiations (see Fig.4, 5 and 6, respectively). For these calculations an enriched ⁶⁷Zn target with the following composition (molar fraction) was supposed: ⁶⁴Zn(\leq 0.5%), ⁶⁶Zn(\geq 9%), ⁶⁷Zn(\geq 80%), ⁶⁸Zn(\leq 10%) and ⁷⁰Zn(\leq 0.5%). The 18 \rightarrow 5, 15 \rightarrow 5 and 12 \rightarrow 5 MeV production energy windows correspond to target thicknesses of 506.8, 362.8 and 226.5 mg/cm², respectively.

In the case of the activation with 12 MeV protons, the available EOB yield is 43.66 MBq/µA (1.18 mCi/µA) and the contamination levels are 3.11% (⁶¹Cu/⁶⁴Cu) and 0.03% (⁶⁷Cu/⁶⁴Cu). Fig.4 shows the time dependence (starting from EOB) of the above activity ratios and the practical ⁶⁴Cu yields. After 7.7 h cooling time, the total contamina-

tion level decays below 1%, but it increases above 1% (due to the longer half life of 67 Cu) after 81.1 h. The available 64 Cu yield (for medical studies) is 28.49 MBq/ μ A (0.77 mCi/ μ A) at the beginning of the application time.




Fig.6. ⁶¹Cu/⁶⁴Cu and ⁶⁷Cu/⁶⁴Cu activity ratio and ⁶⁴Cu activity as a function cooling time using 18 MeV incident proton energy.

15 MeV protons would result 2.3 times higher yield (88.80 MBq/ μ A (2.70 mCi/ μ A) than the 12 MeV route. Contamination levels are 3.74% (61 Cu/ 64 Cu) and 0.04% (67 Cu/ 64 Cu) (see Fig.5.). This process offers a little shorter application window for investigations between 8.9 and 74.4 h after EOB. The useful yield is 61.42 MBq/ μ A (1.66 mCi/ μ A).

In every aspect the bombardment with an 18 MeV beam produces the highest ^{64}Cu EOB yield of 4.22 mCi/µA (156.14 MBq/µA) after 6.3505 h irradiation time (see Fig.6.). Interestingly, this method offers the shortest time range from 9.4 to 70.4 h (after EOB) for applications. The contamination levels are 3.93% ($^{61}\text{Cu}/^{64}\text{Cu}$) and 0.05% ($^{67}\text{Cu}/^{64}\text{Cu}$), while the practical yield is: 93.61 MBq/µA (2.53 mCi/µA) after 9.4 h cooling time.

Longer activation time would result in higher ⁶⁴Cu yields in all cases with decreasing ⁶¹Cu/⁶⁴Cu but with increasing ⁶⁷Cu/⁶⁴Cu activity ratios at EOB, reducing this way the length of the application time.

There are several chemical separation methods in the literature that can be used for separating the radio-coppers effectively from the zinc matrix and the co-formed other radioactive contaminants [29-31]. The knowledge of the activities of these radioisotopes, especially the Gallium nuclides (66 Ga and 67 Ga) at EOB are important not only to select the proper separation technique, but to arrange the handling procedure of the radioactive waste. Based on the available cross section databases of the 66 Zn(p, 2n) 66 Ga, 67 Zn(p, 2n) 66 Ga and 68 Zn(p, 2n) 67 Ga nuclear reactions [19, 22], the 66 Ga yields after 6.3505 h activation are: 196.1, 341.9 and 544.3 MBq/ μ A at 12, 15 and 18 MeV, respectively. For 67 Ga these data are as follows: 284.5 MBq/ μ A (12 MeV), 441.4 MBq/ μ A (15 MeV) and 556.9 MBq/ μ A (18 MeV). The relative high yields of these radio-

gallium isotopes require special care, especially if the recovery of the target material would be necessary within a short time.

4. Conclusions

We have measured the excitation function of the 67 Zn(p, α) 64 Cu nuclear reaction up to 28.5 MeV. Our new data showed good agreement with the updated values of the only data set available in the literature. Since the theoretical calculation (via the TALYS code) predicted different cross sections for the above process, the 64 Cu thick target yields were calculated using the experimental results. Practical production possibilities at three different energies (12, 15 and 18 MeV) were investigated in detail. For all cases highly enriched 67 Zn (>80%) target was suggested with minimum 6.3505 h irradiation time.

It could be concluded that the above reaction can be employed for practical purposes at biomedical cyclotrons, however, it has several drawbacks comparing to the 64 Ni(p, n) 64 Cu process [c.f. 6, 14]. The available yields are not only lower (at 15 MeV around 5 % of the (p, n) route), but it has very large background of 66 Ga and 67 Ga at EOB as well. Additionally, it is more difficult to achieve the required chemical purity of 64 Cu from a 64 Zn target than from a 64 Ni. The high amount of the co-produced Gallium radioisotopes has direct influence in the handling of the irradiated target due to the dose and on the chemical processing.

Based on the calculated yields, the 12 MeV production way fits mainly the requirement for an 'in-house' production (43.66 MBq/ μ A (1.18 mCi/ μ A) at EOB), while the higher energy irradiations (15 and 18 MeV) could provide enough EOB activities (88.80 MBq/ μ A (2.70 mCi/ μ A) and 156.14 MBq/ μ A (4.22 mCi/ μ A), respectively) even for transportation to other medical institutions.

Acknowledgements We thank the operators of the NIRS cyclotron for performing the irradiations. This work was financially supported by the National Institute of Radiological Sciences, (Chiba, Japan) and the Hungarian Research Found (OTKA K108669).

References

- NuDat 2.6 database. Available from: http://www.nndc.bnl.gov/nudat2.
- Qaim S. M., Bisinger, T., Hilgers, K., Nayak, D., Coenen, H. H.: Positron emission intensities in the decay of ⁶⁴Cu, ⁷⁶Br and ¹²⁴I. Radiochim Acta 95, 67-73 (2009).
- 3. Szymański, P., Frączek, T., Markowich, M., Mikiciuk-Olasik, E.: Development of copper based drugs, radiopharmaceuticals and medical materials. Biometals **25**, 1089-1112 (2012).
- Carroll, V., Demoin, D. W., Hoffman, T. J., Jurisson, S. S.: Inorganic chemistry in nuclear imaging and radiotherapy: current and future directions. Radiochim. Acta 100, 653–667 (2012).
- 5. Welch, M. J., Redvanly, C. S. (Eds.), Handbook of Radio-pharmaceuticals. Wiley & Sons, 2003.
- 6. Szelecsényi, F., Blessing, G., Qaim, S. M.: Excitation functions of proton induced nuclear reactions on enriched ⁶¹Ni

- and ⁶⁴Ni: Possibility of production of no-carrier-added ⁶¹Cu and ⁶⁴Cu at a small cyclotron. Appl. Radiat. Isotopes **44**, 575-580 (1993).
- McCarthy, D. W., Shefer, R. E., Klinkowstein, R. E., Bass, L. A., Margeneau, W. H., Cutler, C. S., Anderson, C. J., Welch, M. J.: Efficient production of high specific activity ⁶⁴Cu using a biomedical cyclotron. Nucl. Med. Biology 24, 35-43 (1997).
- 8. Thieme, S., Walther, M., Pietzsch, H.-J., Henniger, J., Preusche, S., Mäding, P., Steinbach J.: Module-assisted preparation of ⁶⁴Cu with high specific activity. Appl. Radiat. Isotopes **70**, 602-608 (2012).
- Zweit, J., Smith, A. M., Downey, S., Sharma, H. L.: Excitation functions for deuteron induced reactions in natural nickel: Production of no-carrier-added ⁶⁴Cu from enriched ⁶⁴Ni targets for positron emission tomography. Appl. Radiat. Isotopes 42, 193-197 (1991).
- Kozempel, J., Abbas, K., Siminelli, F., Zampese, M., Holzwart, U., Gibson, N., Lešetický L.: A novel method for n.c.a. ⁶⁴Cu production by the ⁶⁴Zn(d,2p)⁶⁴Cu reaction and dual ion-exchange column chromatography. Radiochim. Acta. 95, 75-80 (2007).
- 11. Hilgers, K., Stoll, T., Skakun, Y., Coenen, H. H., Qaim S. M.: Cross-section measurements of the nuclear reactions $^{nat}Zn(d,x)^{64}Cu$, $^{66}Zn(d,\alpha)^{64}Cu$ and $^{68}Zn(d,\alpha 2n)^{64}Cu$ for production of ^{64}Cu and technical developments for small-scale production of ^{67}Cu via the $^{70}Zn(p,\alpha)^{67}Cu$ process. Appl. Radiat. Isot. **59**, 343-351 (2003).
- 12. Szelecsényi, F., Steyn, G. F., Kovács, Z., Vermeulen, E., van der Meulen, N. P., Dolley, S. G., van der Walt, T. N., Suzuki, K., Mukai, K.: Investigation of the ⁶⁶Zn(p,2pn)⁶⁴Cu and ⁶⁸Zn(p,x)⁶⁴Cu nuclear processes up to 100 MeV: Production of ⁶⁴Cu. Nucl. Instrum. Meth. **B240**, 625-637 (2005).
- Daraban, L. and Abbas, K. and Simonelli, F., Adam-Rebeles, R., Gibson, N.: Experimental study of excitation functions for the deuteron induced reactions ⁶⁴Zn(d,2p)⁶⁴Cu and ⁶⁴Zn(d,αn)⁶¹Cu using the stacked-foil technique. Appl. Radiat. Isot. 66, 261-264 (2008).
- IAEA Technical Reports Series No. 473.: Nuclear Data for the Production of Therapeutic Radionuclides; (Technical Editors: Qaim, S. M., Tarkanyi, F., R. Capote, R.) ISBN 978-92-0-115010-3 (2011).
- Aslam, M., Sudár, S., Hussain, M., Malik, A. A., Shah, H. A. Qaim S. M.: Charged particle induced reaction cross section data for production of the emerging medically important positron emitter ⁶⁴Cu: A comprehensive evaluation. Radiochim. Acta. 97, 669-687 (2009).
- Experimental Nuclear Reaction Data (EXFOR) (Database Version of February 26, 2013). Available from: www-nds.iaea.org/exfor.
- 17. Levkovskij, V.N.: Activation cross section nuclides of average masses (A=40-100) by protons and alpha-particles with average energies (E=10-50 MeV), Inter Vesi, Moscow, 1991.
- 18. Tárkányi, F., Szelecsényi, F., Kovács, Z., Sudár, S.: Excitation functions of proton induced nuclear reactions on enriched ⁶⁶Zn, ⁶⁷Zn and ⁶⁸Zn. Production of ⁶⁷Ga and ⁶⁶Ga. Radiochim. Acta **50**, 19-26 (1990).
- Szelecsényi, F., Boothe, T. E, Takács, S., Tárkányi, F., Tavano, E.: Evaluated cross section and thick target yield data bases of Zn + p processes for practical applications. Appl. Radiat. Isot. 49, 1005-1032 (1998).
- Koning, A. J., Rochman, D.: Modern nuclear data evaluation with the TALYS code system. Nuclear Data Sheets 113 2841-2934 (2012) or Koning, A., J., Rochman, D., van der Marck, S., Kopecky, J., Sublet, J. Ch., Pomp, S., Sjostrand, H., Forrest, R., Bauge, E., Henriksson, H.: TENDL-2012:

- TALYS-based evaluated nuclear data library, www.talys.eu/tendl-2012.html
- Szelecsényi, F., Kovács, Z., Nagatsu, K., Fukumura K., Suzuki, K., Mukai, K.: Investigation of direct production of ⁶⁸Ga with low energy multiparticle accelerator. Radiochim. Acta 100, 5-11 (2012).
- 22. Gul, K., Hermanne, A., Mustafa, M. G. Nortier, F. M., Oblozinsky, P., Qaim, S. M., Scholten, B., Shubin, Yu., Takács, S., Tárkányi, F. T., Zhuang, Z.: Charged particle cross-section database for medical radioisotope production: diagnostic radioisotopes and monitor reactions. IAEA-TECDOC-1211, IAEA, Vienna, May 2001. Available from: http://www-nds.iaea.org/medical/ (last updated 2011).
- 23. Steyn, G., F., Szelecsényi, F., Kovács, Z., van der Walt, T., N., Dolley, S., G., Vermeulen, C.: Large discrepancies in the excitation function data of the ⁶⁸Zn(p,x)⁶⁴Cu reaction: a possible explanation. J. Phys. Conf. Series 41, 561-564 (2006).
- 24. Anderson, H., H., Ziegler, J., F.: Hydrogen stopping powers and ranges in all elements. In: J.F. Ziegler (Ed.), The stopping and Ranges of Ions in Matter, vol. 3, Pergamon, New York, 1977, p. 1.
- Guide to the expression of uncertainty in measurements, 1995. International Organization for Standardization, Geneva, ISBN 92-67-10188-9.
- Takács, S., Tárkányi, F., Sonck, M., Hermanne, A.: Investigation of the ^{nat}Mo(p,x)^{96mg}Tc nuclear reaction to monitor proton beams: New measurements and consequences on the earlier reported data. Nucl. Instrum. Meth. **B198**, 183-196 (2002).
- 27. Stoll, T., Kastleiner, S., Shubin, Yu., N., Coenen, H., H., Qaim, S., M.: Excitation functions of proton induced reactions on ⁶⁸Zn from threshold up to 71 MeV, with specific reference to the production of ⁶⁷Cu. Radiochim. Acta **90**, 309-313, (2002).
- 28. Kastleiner, S., Coenen, H., H., Qaim, S., M.: Possibility of production of ⁶⁷Cu at a small-sized cyclotron via the (p,)-reaction on enriched ⁷⁰Zn. Radiochim. Acta **84**, 107-100 (1999).
- Dirks, C., Scholten, B., Happel, S., Zulauf, A., Bombard, A., Jungclas, H.: Characterisation of a Cu selective resin and its application to the production of ⁶⁴Cu. J. Radioanal. Nucl. Chem, 286, 671-674 (2010).
- Al Rayyes, A. H., Ailouti, Y.: Routine simultaneous production of no-carrier-added high purity ⁶⁴Cu and ⁶⁷Ga. Nukleonika 56, 259-262 (2011).
- Asad, A. H., Smith, S. V., Chan, S., Jeffery, C. M., Morandeau, L., Price, R. I.: Cyclotron production of ⁶¹Cu using natural Zn & enriched ⁶⁴Zn targets. AIP Conf. Proc. **1509**, 91-95 (2012).